


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-9C-C-N-QTKA

S.N.: 10222

Date: JUN 8 2002



- a) Output signal amplitude:
0.70 - 1.00 mA
- b) Pulse width:
0.5 TO 10 NS
- c) Rise time:
≤ 300 ps
- d) Fall time:
≤ 600 ps.
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 to 1 Hz.

TOP MAIN PLANE OUT
TO 40 dB.

BUT AX-51 MI OUT TO 20 dB
IN 451A 4500 MS A TEST LOAD

PRF = 12 MHz.