


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANPP-1-C-PN-60*

S.N.: *10223*

Date: *MAY 27 2002*



- a) Output signal amplitude:
0 to ± 10 Volts (to 50V)
- b) Pulse width:
0.2 to 100 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 to 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V

40 dB ATTEN: 10 V/DIV
TOP 0.2 to 5.0 ns RANGE
500 ps/DIV

BOT 5 to 100 ns RANGE
5 ns/DIV

PRF = 100 KHz

50-160 MHz

WA