


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV1R-1-P*

S.N.: *10226*

Date: *JUNE 04 2002*



- a) Output signal amplitude:  
*0 TO +200 V (7050A)*
- b) Pulse width:  
*2 TO 5 NS*
- c) Rise time:  
*≤ 0.9 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF: *0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*60 DB ATTN*

*-200 VOLTS/DIV*

*PRF ≈ 10 KHZ*

*+15 VOLTS DC*  
*600 mA (MAX)*