


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-3-C-PN-00*

S.N.: *10231*

Date: *June 12 2002*



- a) Output signal amplitude: *0 to ±10 Volts*
- b) Pulse width: *1 to 5 ns*
- c) Rise time: *≤ 300 ps*
- d) Fall time: *≤ 600 ps*
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 Hz

40 dB ATTN

10 Volts/div

TOP 1 ns/div, PRF = 10 MHz

BOT 5 ns/div, PRF = 25 MHz

[Signature]