


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-53-C-P-QTKA

S.N.: 10274

Date: JUN 30 2002



- a) Output signal amplitude: 0 to +40V (70-50%)
- b) Pulse width: 1.5 to 10ns
- c) Rise time: 0.5 ns (20-80%)
- d) Fall time: ≤ 1 ns
- e) PRF: 0 to 1 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V

50 dB ATTN
- 32 VOLTS/DIV
2 ns/DIV

PRF = 100 kHz

2

50-60 Hz