


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-P-DN-BA-SW-ED*

S.N.: *10277*

Date: *JUL 12 2002*



- a) Output signal amplitude:
0.70 ± 500mV (70.50mV)
- b) Pulse width:
2.270 (6N)
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 135ps
- e) PRF:
0.70 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

40 dB ATTN

0.5 V/DIV

TOP 1ns/DIV

BOI 100ps/DIV
(RISE TIME)

PRF = 12 MHz