


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-N*

S.N.: *10279*

Date: *JUNE 6 2002*



- a) Output signal amplitude:
0 TO -5V (TO 50%)
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTN
5 VOLT/DIV
PRF = 10 MHz

120/240 V
50-60 MHz
[Signature]