


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANL-2A-PS-CM1-05-150-P

S.N.: 10298

Date: JULY 23 2002



- a) Output signal amplitude:
0 TO +170 VOLTS (TO 50n)
(0 TO +150 AT PRF = 20 KHz)
- b) Pulse width:
3 TO 100 nS
- c) Rise time:
≤ 2 nS
- d) Fall time:
≤ 3 nS
- e) PRF: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz.

60 dB ATTN

-100 VOLT/DIV

PRF = 10 KHz