


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AN-2-P*

S.N.: *10316*

Date: *JULY 18 2002*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width: *(TO 50ns)*
2 TO 50 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:

TR 5 NS/DIV

VMID 200 ps / DIV (RISE TIME)

VF 20 ps / DIV (FALL TIME)

40 dB ATTN (-10 VOLTS / DIV)

PRF = 50 kHz

+15 VDC
400 mA (max)

[Signature]