


PULSE GENERATOR
PERFORMANCE CHECK

Model: *NH-5-1-P*
 S.N.: *10332*
 Date: *SEPT 12 2002*



- a) Output signal amplitude: *0 to +10 V*
- b) Pulse width: *0 to 50 ns*
≤ 130 ps
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*
250 mA (max)

40 dB ATTEN
10 VOLT/DIV
TOP 200ps/DIV
Bot 1 ns/DIV
PRF = 100KHz

[Signature]