


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AV-1402-C-P*

S.N.: *10337*

Date: *SEPT 6 2002*



- a) Output signal amplitude:  
*0 TO +20 V (70.500V)*
- b) Pulse width:  
*300ps TO 2.0NS*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 200 ps*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*

*40 dB ATTN*  
*10 V/DIV*  
*PRF = 100 kHz*