


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AN-1-P*
S.N.: *10345*
Date: *SEPT 12 2002*



- a) Output signal amplitude:
0 TO +10 VOLTS (7050V)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC
250 mA (MAX)

40 dB ATTEN
- 10 VOLTS/DIV
TOP 1 NS / DIV (PW MAX)
PROT. 200 ps / DIV (PW MIN)
PRF ≈ 100 KHz