


PULSE GENERATOR
PERFORMANCE CHECK

Model: ANM4-1-C-P-M

S.N.: 10347

Date: SEPT 27 2002



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width: (TO 50ns)
≤ 200 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 100 ps
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V

40 dB ATTN

10 VOLTS/DIV

PRF = 12 MHz.

50-60 Hz