


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-1-e-N-M*

S.N.: *10348*

Date: *SEPT 27 2002*



- a) Output signal amplitude:
0 to -10 Volts
- b) Pulse width: *(to 50%)*
≤ 200ps
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 100ps
- e) PRF:
0 to 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

40 dB ATTEN
-10 Volts/DIV
PRF = 12 MHz

120/240 V
50-60 Hz
[Signature]