


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AMP-AV-143-C-P-PN

S.N.: 10362

Date: Aug 28 2002



- a) Output signal amplitude:  
0 to ± 20 volts
- b) Pulse width: (to 50ns)  
0.4 to 2.0 ns
- c) Rise time:  
≤ 150 ps
- d) Fall time:  
≤ 250 ps
- e) PRF: 0 to 100 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V  
50-60 Hz

50 dB ATTN  
32 VOLTS/DIV.  
TOP MAX PW  
BOI MIN PW  
PRF = 100 kHz.