

PULSE GENERATOR
PERFORMANCE CHECK

Model: *ACRH-3. PS PN*

S.N.: *10372*

Date: *OCT 2 2002*

- a) Output signal amplitude:
0 TO $\pm 3000V$ ($R_L \geq 100\Omega$)
- b) Pulse width:
200NS TO 2.5 μ S
- c) Rise time:
 $\leq 100NS$
- d) Fall time:
 $\leq 100NS$
- e) PRF:
0 TO 1KHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz.



10372

$R_c = 1K.$

NEG output

PRF = 100 Hz

