


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVRL-1-PS-M

S.N.: 10396

Date: OCT 7 2002



- a) Output signal amplitude:  
-200 VOLTS (TO 50M)
- b) Pulse width:  
5 TO 100 NS
- c) Rise time:  
≤ 2 NS
- d) Fall time:  
≤ 3 NS
- e) PRF: 0 TO 5 KHZ.
- f) Jitter, stability:  
OK
- g) Prime power:  
100 ± 10 VOLTS  
50\* - 60 HZ.

60dB ATTEN  
-100 VOLTS / DIV  
TOP 10 NS / DIV  
BOT 2 NS / DIV (RISE TIME)  
PRF = 2 KHZ.