


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1A3-PS-M-P*

S.N.: *10407*

Date: *OCT 24 2002*



- a) Output signal amplitude:
0.70 + 40 VETS
- b) Pulse width: ~~(to 50 ns)~~
0.4 to 2.0 NS
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 250 ps
- e) PRF:
0.70 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz.

50 dB ATTN

32 VOLT/DIV

TOP PW OFF MAX

BOT PW ON MIN

PRF = 100K Hz

[Signature]