


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-3-PR1*
S.N.: *10420*
Date: *SEPT 27 2002*



- a) Output signal amplitude: *+100 TO +1500V*
- b) Pulse width: *(TO 50ns)*
≤ 5 NS (AT THE POWER)
- c) Rise time: *≈ 1 NS*
- d) Fall time: *≤ 10 NS*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:

60 dB ATTN
- 100 VOLTS/DIV
PRF = 10 KHz

+15 VDC, 600mA
MAX

[Signature]