


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR14-2-C-P  
S.N.: 10443  
Date: DEC 6 2002

- a) Output signal amplitude:  
0 to +2000 V
  - b) Pulse width: CR<sub>2</sub> 10K  
0.25 to 2.5 μs
  - c) Rise time:  
≤ 80 ns (20-80%)
  - d) Fall time:  
≤ 80 ns (20-80%)
  - e) PRF: 0 to 1 kHz
  - f) Jitter, stability:  
OK
  - g) Prime power:  
120/240 V  
50-60 Hz
- 

10443

$R_L \approx 10K$

$PRF \approx 1KHz$

