

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-112A-PS R5*
S.N.: *10444*
Date: *DEC 12 2002*


- a) Output signal amplitude:
0 TO ± 200 V
CR_L ≥ 5 A
- b) GAIN: ...
x 1 TO x 100
- c) Rise time:
≤ 1 μs
- d) Fall time:
≤ 1 μs
- e) PRF:
DC TO 300 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

(A)

10444

$R_L = 6K$


$C_L = 0 \mu fd$



10444

$$R_L = 6K.$$

$$C_L = 3000 \mu\text{fd.}$$



©

10444

$R_c = 6K$

$C_L = 0pfd$

$f = 300 KHz$

