


PULSE GENERATOR  
PERFORMANCE CHECK

Model: ANM4-4-CP  
S.N.: 10458  
Date: DEC 19 2002



- a) Output signal amplitude: 0 to +100 V ✓
- b) Pulse width: (to 50%)  $\leq 1.0 \text{ ns}$
- c) Rise time:  $\text{N.S.}$
- d) Fall time:  $\text{N.S.}$
- e) PRF: 0 to 10 MHz
- f) Jitter, stability:  $\text{OK}$
- g) Prime power: 120/240V

60 dB ATTN  
- 100 V/240V/DN  
PRF = 1.0 MHz

50-60 MHz  
*[Signature]*