


PULSE GENERATOR
PERFORMANCE CHECK

Model: MM-1-C-N
S.N.: 10467
Date: JAN 06 2003



- a) Output signal amplitude: 0.70 - 5V (70 50V)
- b) Pulse width: 0.270 6.0 ns
- c) Rise time: ≤ 100ps
- d) Fall time: ≤ 135ps
- e) PRF: 0.70 25 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/240V

40 dB ATTN ∴ 5V/DIV

50 * 60 MHz

TOP 1 ns/DIV

BOI 100 ps/DIV (LOADING
EDGE RISE TIME)

PRF = 12 MHz