


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMH-5-C-P
S.N.: 10488
Date: FEB 18 2003



- a) Output signal amplitude: 0 TO +100 V
- b) Pulse width: (70 50%)
2 TO 4 NS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power: 120 / 240 V
50 - 60 Hz

60 dB ATTN
- 100 VOLTS / DIV
TUP PW MIN
POT. PW MAX
PRF = 100 kHz

[Signature]