


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-7011.B1-B
 S.N.: 10503
 Date: FEB 25 2003



60 dB ATTN.: 100 VOLTS/DIV


- ① a) Output signal amplitude:
0 TO ± 100 VOLTS (70.50 V)
- ② b) Pulse width:
100 NS TO 1 μS
(5% MAX DUTY CYCLE)
- ③ c) Rise time:
≤ 2 NS
- ④ d) Fall time:
≤ 2 NS
- e) PRF: 0 TO 100 KHz
(5% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

- ① Point 5 ns/div, RISE TIME (20-80%)
- ② Point 5 ns/div, FALL TIME
- ③ Point 5 ns/div, RISE TIME
- ④ Point 5 ns/div, FALL TIME

PRF = 10 KHz
 PW ≈ 900 NS

(A)

10503
WIDE PULSE
 $R_L = 50\Omega$
PRF = 100 Hz
POSITIVE OUT



③


10503

WIDE PULSE

$R_L = 50\Omega$

PRF = 100 ITB


NEGATIVE OUT.



©


10503

VERY WIDE PULSE,
RISE TIME
 $R_c = 50 \Omega$
 $PW = 250 \mu s$
POSITIVE OUT.



← LIMITED
BY 300MHz
SCOPE.

25 Feb 2003
16:23:47

AVTECH**AVTECH ELECTROSYSTEMS LTD.**NANOSECOND WAVEFORM ELECTRONICS
SINCE 1975
 P.O. BOX 265
 OGDENSBURG, NY
 U.S.A. 13669-0265
 TEL: (315) 472-5270
 FAX: (613) 226-2802

 TEL: 1-800-265-6681
 FAX: 1-800-561-1970

 e-mail: info@avtechpulse.com
 http://www.avtechpulse.com

 BOX 5120, LCD MERIVALE
 OTTAWA, ONTARIO
 CANADA K2C 3H4
 TEL: (613) 226-5772
 FAX: (613) 226-2802
"-B" Functional Test & Calibration Certificate

Date of test:	February 25, 2003				Tester:	MJC
Programmed model name:	AV-1011B1-B					
Programmed serial number:	10503					
Firmware revision:	2.44					
Internal trigger checked at:	1 Hz	10 Hz	1 kHz	10 kHz	100 kHz	
Actual measured output ¹ :	1.001 Hz	99.76 Hz	0.998 kHz	9.976 kHz	99.75 kHz	
External trigger checked:	yes			Gate checked:	yes	
Manual trigger checked:	yes					
Pulse compression checked:	yes	Low Amplitude PW Distortion Nulled:			N/A	
Pulse width checked at:	100 ns	1 us	10 us	1 ms	10 Hz, +100V	
Actual measured output ² :	99.0 ns	1.007 us	10.2 us	1.058 ms	to 50 Ohms	
PWin = PWout mode checked:	yes	DC mode checked:			N/A	
Duty Cycle Limit:	5%					
Delay nulled:	yes					
Delay checked at:	100 ns	1 us	10 us	1 ms	10 Hz, +100V	
Actual measured output ¹ :	100.0 ns	0.999 us	9.995 us	1.005 ms	to 50 Ohms	
Double pulse checked:	yes					
Invert mode checked:	N/A					
ECL/TTL modes checked:	N/A					
Zout switch checked:	yes					
Amplitude checked at:	-10V	+20V	-50V	+100V	100 Hz, 100 us	
Actual measured output ² :	-10.1V	+19.9V	-50.2V	+100.0V	into 50 Ohms	
Amplitude polarity:	+/-					
Zout calibration:	N/A					
Electronic amplitude control:	OK					
External amplify mode:	N/A					
Bleeder resistors adequate:	yes					
Ultraviolet flux removed:	yes					
Monitor V/I Ratio:	N/A			Monitor offset nulled:		
LCD Monitor calibrated:	N/A					
Offset checked at:	N/A					
Actual measured output ² :	N/A					
Offset nulled (output on):	N/A			Amplitude-dependent offset nulled:		
Offset nulled (output off):	N/A					
RS-232 checked:	yes					
LCD pull-ups installed:	yes					
PN trigger pull-downs installed:	N/A					
PW stable during amplitude changes:	yes					
Sync pulse width checked:	50 ns nom.					
Circuit Boards:	PS:	93	Main:	108B		
Overload Trigger Resistance:	Trips at:	N/A	Installed:	3.9k		
DC fuses:	Positive:	2.5A, 1.6A	Negative:	N/A		
AC Current:	Quiescent:	0.18A@230V 0.29A@115V	Max. Load:	0.29A@230V 0.55A@115V		
AC fuse:	two 0.5A, for 240V use					
120/240V operation:	OK			Fan operational:	yes	
Photographed:	yes					

¹ Checked with: Fluke PM6681 Counter (S/N 9446 066 81016),
 referenced to Datum ExacTime 9390-6000 (S/N 4461) GPS Frequency Reference

² Checked with: Tektronix TDS3052 digital oscilloscope (S/N B014783) for PW ≥ 5 ns,
 Tektronix 7704A/7S11/7T11/S4 sampling oscilloscope system (Cal. Label 112506) for PW < 5 ns.