


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-3AH-C-P*

S.N.: *10525*

Date: *DEC 17 2002*



- a) Output signal amplitude:  
*0 TO +450 V (TO 50W)*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 1 NS*
- d) Fall time:  
*± 4 NS*
- e) PRF:  
*0 TO 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 Hz*

*60 dB ATTEN ∴ 100V/DIV*

*TOP 10 NS/DIV*

*BOE 1 NS/DIV (RISE TIME)*

*PRF = 2.0 kHz*