

PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-3-C-N-KOPA-EA

S.N.: 10527

Date: FEB 19 2003

- a) Output signal amplitude:
0 TO -300V (TO $R_L \geq 50\Omega$)
- b) Pulse width:
100 NS TO 1.0 μ S
- c) Rise time:
 ≤ 15 NS
- d) Fall time:
 ≤ 15 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50* - 60 Hz



(A)

10527

BOTH outputs $R_L = 50\Omega$
PBF = 1 KHz

