


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MX-3-C-N-KOPA-CA*

S.N.: *10530*

Date: *FEB 27 2003*

- a) Output signal amplitude:  
*0 TO 300V (TO 1.75μs)*
- b) Pulse width:  
*100NS TO 1.005*
- c) Rise time:  
*≤ 15 NS*
- d) Fall time:  
*≤ 15 NS*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-160 Hz*



10530

BOTH OUTPUTS  $R_L = 50\Omega$

DUTY = 1 KHz.

