



P.O. BOX 265  
OGDENSBURG, NY  
U.S.A. 13669-0265

TEL: 888-670-8729 (USA & Canada) or +1-613-686-6675 (Intl)  
FAX: 800-561-1970 (USA & Canada) or +1-613-686-6679 (Intl)

BOX 5120, LCD MERIVALE  
OTTAWA, ONTARIO  
CANADA K2C 3H5

info@avtechpulse.com - http://www.avtechpulse.com/


PERFORMANCE CHECKSHEET

Model: AVR-3-PW-C-P-TEK2  
Type: High-Speed Pulse Generator  
S.N.: 10541 (repair)  
Date: February 20, 2019

Output Amplitude: up to +250V, to 50Ω  
Pulse Width (FWHM): 0.25 – 250 us  
Rise Time (20%-80%): ≤ 10 ns  
Fall Time (80%-20%): ≤ 10 ns  
PRF: 1 Hz - 10 kHz  
Jitter, Stability: OK  
Prime Power: 120/240V AC, 50-60 Hz.


Basic specifications: →

Narrow Pulse, Voltage Waveform, No Shorting Bar




Output of "OUT" connector, terminated into the internal 50 Ohm load using a 12" length of RG-316 cable (i.e., no shorting bar). Viewed with TDS3052B scope. 50V/div, 400 ns/div. 10 Hz.

Wide Pulse, Voltage Waveform, No Shorting Bar




Output of "OUT" connector, terminated into the internal 50 Ohm load using a 12" length of RG-316 cable (i.e., no shorting bar). Viewed with TDS3052B scope. 50V/div, 40 us/div. 10 Hz.

Narrow Pulse to Shorting Bar and A6303 Probe



Output of A6303 probe and AM503B amplifier, viewed with TDS3052B scope. The A6303 probe is clamped to the user-supplied shorting cable (not the original shorting bar).

Wide Pulse to Shorting Bar and A6303 Probe



Output of A6303 / AM503B, viewed with TDS3052B scope. The probe is clamped to the user-supplied shorting cable (not the original shorting bar).

References levels: 20%, 80%