


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-M-HU2-N-M-EW*

S.N.: *10571*

Date: *MAR 20 2003*



- a) Output signal amplitude:
0 to -20 V (to 50V)
- b) Pulse width:
0.370 2.0 ns
- c) Rise time:
≤ 100 ns
- d) Fall time:
≤ 200 ps
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
300 mA (MAX)

40 dB ATTN
10 VOLTS/DIV

TR 1 NS/DIV

PROV 100 ps/DIV
(RISE TIME)

PRF = 100 kHz

OK

[Signature]