


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV1M14-4-C-O-M-P-PN

S.N.: 10621

Date: MAY 12 2003



- a) Output signal amplitude:
070 ± 100 V (70.50V)
- b) Pulse width:
≤ 1 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: 070 10 MHz.
- f) Jitter, stability: OK
- g) Prime power: 170/240 V
5.0 - 60 Hz.

60 dB ATTN

100 VOLTS/DIV

PRF = 10 MHz