


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-N-1752-N-MA-EW*

S.N.: *10635*

Date: *SEPT 2 2003*



- a) Output signal amplitude:  
*0 TO -20V (TO 50N)*
- b) Pulse width:  
*0.3 TO 2.0NS*
- c) Rise time:  
*≤ 100ps*
- d) Fall time:  
*≤ 200ps*
- e) PRF:  
*0 TO 1MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VDC*  
*300 mA (MAX)*

*40 dB ATTEN*  
*-10 V/DIV*

*PRF = 100 KHz.*

*[Signature]*