


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMH-3-C-P-PN-M-05*

S.N.: *10681*

Date: *SEPT 17 2003*



- a) Output signal amplitude:
0 to ± 50 VOLT. (250Ω)
- b) Pulse width:
≤ 600ps
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 to 25 MHz.*
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

50 dB ATTN

32 VOLT/DIV

PRF = 12 MHz.