


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-MU-1-C-N-05-F*

S.N.: *10705*

Date: *OCT 29 2003*



- a) Output signal amplitude: *0 TO +10 VOLTS.*
- b) Pulse width: *(TO 50ns)*
0.2 TO 4.0 NS
- c) Rise time: $\leq 100ps$
- d) Fall time: $\leq 135ps$.
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

120 - 240 V
50 - 60 Hz.

40 dB ATTN. ∴ 10V/DIV

TOP 1 NS/DIV

BOT 100ps/DIV (RISE TIME)

PRF = 100 kHz.

M
