


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-5-C-F*

S.N.: *10711*

Date: *OCT 7 2003*



- a) Output signal amplitude:
0 TO +100V (TO 50Ω)
- b) Pulse width:
2 TO 4NS
- c) Rise time:
≈ 1 NS
- d) Fall time:
≈ 2 NS
- e) PRF:
0 TO 1 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

60 LB ATTEN

100 V/V B/DIV

PRF = 100 KHz

[Signature]