

PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVIR-1-C-P-AT-PWT

S.N.: 10721

Date: OCT 16 2003



- a) Output signal amplitude: 0 to +200V to 50 $\Omega$
- b) Pulse width: 2 to 5 ns
- c) Rise time:  $\leq 0.8$  ns
- d) Fall time:  $\leq 1$  ns
- e) PRF: 0 to 50 kHz
- f) Jitter, stability:

60 dB ATTN.: 100V/DIV

TOP MAX PW

BOT MIN PW

PRF  $\approx$  20 kHz

- g) Prime power: OK  
120/240V  
50-60 Hz.