


PULSE GENERATOR
PERFORMANCE CHECK

Model: PNR-64-C-P-SMA-ED

S.N.: 10764

Date: NOV 4 2003



60dB ATTN. - 100V/DIV

TOP 5 ns/DIV
(MAX PW)

BOT 1 ns/DIV
(MIN PW)

PRF = 100KHz

- a) Output signal amplitude:
0 TO +100 V (TO 50n)
- b) Pulse width:
1 TO 5NS
- c) Rise time:
≤ 0.4 NS
- d) Fall time:
≤ 0.4 NS
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
100 TO 240V
50 TO 60 Hz