


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-P-6A-5W*

S.N.: *10785*

Date: *NOV 18 2003*



- a) Output signal amplitude:  
*0 TO +5V (TO 50W)*
- b) Pulse width:  
*0.2 TO 6.0 NS*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 135 ps*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*40 dB ATTN ∴ 5 V/DIV*

*TOP 1 NS/DIV*

*BOT 100 ps/DIV (RISE TIME)*

*PRF = 10 MHz*

*+24 VDC*  
*500mA (MAX)*