


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-15-C-N

S.N.: 10811

Date: JAN 05 2004



- a) Output signal amplitude:
0 to 5V (to 50Ω)
- b) Pulse width:
0.2 to 4.0ns
- c) Rise time:
≤ 60 ps.
- d) Fall time:
≤ 135 ps
- e) PRF:
0 to 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V

50-60 MHz
[Signature]

40dB ATTN
5 V/DIV
TOP 500ps/DIV
BOT. 50ps/DIV (RISE TIME)
PRF = 1.0 MHz.