


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PVM-1-C-P*

S.N.: *10838*

Date: *FEB 3 2004*



- a) Output signal amplitude:
0 TO +5V (TO 50mV)
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
100 → 240 V
50 → 60 Hz

40 dB ATTN. ∴ 5 V/DIV

- 1) *20 NS/DIV PRF = 12 MHz*
- 2) *1 NS/DIV " " "*
- 3) *100 ps/DIV (RISE TIME)*