


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-3A-C-P-M-05*

S.N.: *10839*

Date: *FEB 13 2004*



- a) Output signal amplitude:
0 TO +300V (to 50Ω)
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 600ps (20-80%)
- d) Fall time:
≤ 4 NS
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability: *OK*
- g) Prime power:
120 / 240V
50-60Hz.

60 dB ATTN.: 100V/DIV
POAF ≈ 3 KHz.

TOP 10 NS/DIV

BOT. 1.0 NS/DIV (RISE TIME)