


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-C-OT-P*

S.N.: *10870*

Date: *MAR 12 2004*



- a) Output signal amplitude:  
*0 TO +5V TO 50Ω*
- b) Pulse width:  
*0.2 TO 6.0ns*
- c) Rise time:  
*≤ 100ps*
- d) Fall time:  
*≤ 135 ps.*
- e) PRF:  
*0 TO 25 MHz.*
- f) Jitter, stability:
- g) Prime power: *OK*

*40 dB ATTEN*

*5 VOLT/DIV*

*1ns/DIV*

*100ps/DIV*

*(RISE TIME)*

*100 → 240V*  
*50-60Hz.*