


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM4-4-C-7-MSHA*

S.N.: *10881*

Date: *APRIL 5 2004*



- a) Output signal amplitude: *0 to +100V (to 50V)*
- b) Pulse width: *≤ 1ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *100 → 240V*
50 - 60 Hz

60dB ATTN. ∴ 100V/DIV

PRF ≈ 1.0 MHz

[Signature]