


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-P*

S.N.: *10897*

Date: *APRIL 20 2004*



- a) Output signal amplitude:
0 to +5V (to 50Ω)
- b) Pulse width:
200ps to 6.0ns
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 135ps
- e) PRF:
0 to 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*40dB ATTN.: 5V/DIV
Rise 1ns/DIV
PW 200ps MAX*

*PrI 500ps/DIV
PW 200ps MIN
PRF = 10 MHz*

*+24V DC,
500mA (MAX)*

