


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-NIM-ED-BU-6A-M-P*

S.N.: *10933*

Date: *MAY 20 2004*



40 dB ATTN.: 5 VOLTS/DIV

TOP: 1 NS/DIV

BOT: 100 PS/DIV

PRF = 10 MHz

- a) Output signal amplitude:
0 TO 75 V (to 50 mA)
- b) Pulse width:
0.2 TO 6 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+ 24V DC
500 mA MAX

A large, stylized handwritten signature in black ink, located at the bottom right of the page.