


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-141-P

S.N.: 10935

Date: JUNE 4 2004



- a) Output signal amplitude:  
0 TO +100 V (to 50Ω)
- b) Pulse width:  
( $\leq 1.0$  NS)
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF: 0 TO 100 KHz.
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VDC  
250mA MAX

60dB ATTN

100 V/DIV

PRF = 100KHz.