


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-2-C-P-ECL*

S.N.: *10945*

Date: *JULY 6 2004*



- a) Output signal amplitude:
0 to +15V (to 50Ω)
- b) Pulse width:
0.2 to 2.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 35 ps
- e) PRF:
0 to 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

100 → 240 V
50 → 60 Hz

40 dB ATTEN
-10 V/DIV.

700 1 ns/DIV

PROT 100ps/DIV (RISE TIME)