


PULSE GENERATOR  
PERFORMANCE CHECK

Model: ANM4-3-C-P-05  
S.N.: 10947  
Date: AUG 09 2004



- a) Output signal amplitude:  
0 to +50V (7050W)
- b) Pulse width:  
≤ 600ps
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 to 25 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

50 dB ATTEN  
∴ .32 VOLTS/DIV  
TOP 500 ps/DIV  
BOT. 5 ns/DIV  
PRF ≈ 12 MHz

100 → 280V  
50 → 60 MHz  
