


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-101-TR-C-P*  
 S.N.: *10967*  
 Date: *AUG 30 2004*



- a) Output signal amplitude:  
*≥ +50 V (70.50V)*
- b) Pulse width:  
*2 TO 100 NS*
- c) Rise time:  
*≤ 200 ps*
- d) Fall time:  
*≤ 2 ns*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 dB ATTN*  
*32V/DIV*

*5 ns/DIV*

*200 ps/DIV*  
*(RISE TIME)*

*PRF = 20 KHz*

*100 → 240 V*  
*50 → 60 Hz*