


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVMM-3-C-P-PN-M-D-TR-0T

S.N.: 10973

Date: SEPT 3 2007



- a) Output signal amplitude:  
0 TO  $\pm 10V$  ( $\approx 50\Omega$ )
- b) Pulse width:  
1 TO 5 NS
- c) Rise time:  
0.3 TO 2.0 NS
- d) Fall time:  
0.6 TO 2.0 NS
- e) PRF:  
0 TO 25 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
100  $\rightarrow$  240V  
50 - 60 MHz

40 dB ATTN.: 10V/DIV  
RISE TIME ON ALL  
5 SETTINGS  
PRF = 12 MHz.

100  $\rightarrow$  240V  
50 - 60 MHz  
AM